

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE**CERTIFICAT D'ESSAI OC**

Product
Produit

Power supply for building-in

Name and address of the applicant
Nom et adresse du demandeur

EGSTON System Electronics Eggenburg GmbH
Grafenberger Strasse 37, A-3730 Eggenburg, Austria

Name and address of the manufacturer
Nom et adresse du fabricant

EGSTON System Electronics Eggenburg GmbH
Grafenberger Strasse 37, A-3730 Eggenburg, Austria

Name and address of the factory
Nom et adresse de l'usine

1) EGSTON System Electronic spol.s.r.o.
Průmyslova 20, CZ-670 15 Znojmo, Czech Republic

Note: When more than one factory, please report on page 2
Note: Lorsque il y a plus d'une usine, veuillez utiliser la 2^{ème} page

Additional Information on page 2

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

Input: 100-240 Vac; 50-60 Hz; 2,0 A
Output: See page 2

Trademark (if any)
Marque de fabrique (si elle existe)

EGSTON

Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais constructeur

/

Model / Type Ref.
Ref. De type

E2OFxW3 120

Additional information (if necessary may also be reported
on page 2)

The risk management requirements of the standard were not
addressed.

Les informations complémentaires (si nécessaire, peuvent
être indiqués sur la 2^{ème} page

Additional Information on page 2

A sample of the product was tested and found
to be in conformity with
Un échantillon de ce produit a été essayé et a été
considéré conforme à la

IEC 60601-1:2005 (3rd Ed.) + A1:2012

As shown in the Test Report Ref. No. which forms
part of this Certificate
Comme indiqué dans le Rapport d'essais numéro de
référence qui constitue partie de ce Certificat

T223-0473/15, dated 2015-12-21


This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



SIQ Ljubljana
Tržaška cesta 2, SI-1000 Ljubljana, Slovenia
T +386 1 4778 100, F +386 1 4778 444, info@siq.si, www.siq.si
Product Certification Body is accredited by Slovenian Accreditation, Reg. No.: CP-001

Date: 2015-12-21

Signature: Alja Pregl



Name and address of the factory:

- 1) EGSTON System Electronic spol.s r.o.
Průmyslová 20, CZ-670 15 Znojmo, Czech Republic
- 2) EGSTON Electronics Zhuhai Ltd.
3/F, No. 3 Pinggong No. 1 Road, Nanping Technology Ind. Park, Zhuhai City, PR China

Ratings and principal characteristics:

Input:

100-240 Vac; 50-60 Hz; 2,0 A

Output:

Output voltage: 12 Vdc

Output current: 8,3 A at 100 Vac voltage

Output current: 10,0 A at 240 Vac voltage

Output voltage: 24 Vdc

Output current: 4,2 A at 100 Vac voltage

Output current: 5,0 A at 240 Vac voltage

Additional information (if necessary)**Information complémentaire (si nécessaire)**

Date: 2015-12-21

Signature: Alja Pregl

